

FDA Approved Home COVID-19 Test Kits

Home Test	Extracting Reagent ingredients, volume (if available)	Product warnings	Storage container
BinaxNOW (Abbott)	Sodium azide 0.025%, 1mL	Nonspecific	Dropper
QuickVue (Quidel)	Sodium phosphate Dibasic 0.7%, C12-14-alkyldimethylbetaines 0.7%), ProClin™ 300, EDTA 0.03%	Eye irritation/damage, burns, allergic reactions	Test-tube
InteliSwab (OraSure)	Triton™ X-100 (detergent), ProClin™ 950, 0.75mL	Non-toxic	Test-tube
BD Veritor (BD)	0.1% sodium azide, Triton™ X-100, 0.325mL	Harmful if inhaled, swallowed, or exposed to skin	Test-tube
Ellume	ProClin® 300	Allergic skin reactions	Dropper
Flowflex (ACON)	Sodium Azide 0.02%, Triton™ X-100	Hazardous to skin/eye	Test-tube
CareStart (Access Bio)	Sodium tetraborate, ETDA, Triton™ X-100, N-lauroylsarcosine, 0.5mL	Eye irritant/damage, harmful if swallowed, skin irritation, toxic to aquatic life	Test-tube
iHealth	Triton™ X-100 0.1%, ProClin™ 300 0.05%	Skin/eye irritation, burns, allergic reactions	Test-tube
SCoV-2 Ag Detect (InBios)	Tris(hydroxymethyl)aminomethane, Nonidet P-40, sodium dodecyl sulfate, sodium deoxycholate, EDTA (all <0.01%)	Skin/eye irritation, toxic to aquatic life	Dropper
Celltrion DiaTrust (Celltrion USA)	Sodium azide, 0.09% 0.3mL	Skin/eye irritation/damage	Test-tube

Source: In Vitro Diagnostics EUAs - Antigen Diagnostic Tests for SARS-CoV-2. U.S. Food and Drug Administration. December 29, 2021. Accessed Jan 3, 2022. <https://www.fda.gov/medical-devices/coronavirus-disease-2019-covid-19-emergency-use-authorizations-medical-devices/in-vitro-diagnostics-euas-antigen-diagnostic-tests-sars-cov-2>